AID IN COMPLETING APPLICATION FOR RADIOACTIVE MATERIAL LICENSE - LABORATORY/INDUSTRIAL

For the most part the form is self-explanatory. Fill in the blanks, check the appropriate boxes, and attach the additional information as requested. Either use standard application appendices, where available, or submit equivalent procedures. These should be numbered according to the corresponding item number in the application.

Please submit the application and all attachments, with the exception of Regulatory Guide 10.7, which is for your information only.

Checking your work:

- ÿ Have you included the street address of the facility?
- $\ddot{\mathsf{y}}$ Are you sure the sealed sources (if any) match the exposure devices according to manufacturer's data?
- \(\bar{Y} \) Have you completed and included all the requested attachments?

Attachment A for RSO and each user - Training and Experience

Attachment B or equivalent - Duties of the Radiation Safety Officer

Attachment C or equivalent procedures - Calibration of Exposure Rate Instruments

Pocket dosimeter check methods (unless approved service agency used)

Detailed bioassay program

Facilities and storage diagram

Facilities and equipment description

Attachment D or equivalent - Standard Survey Criteria and Procedures

Detailed leak test counting procedures (unless approved service agency used)

Attachment E - Records Required

Sample of each record form

Attachment F or equivalent instructions - Personnel Training Program

Attachment G or equivalent lab safety rules - General Rules for Safe Use

Attachment H or equivalent procedures - Waste Disposal

Attachment I or equivalent procedures - Emergency Procedures

Attachment J or equivalent procedures - Ordering and Accepting Delivery

Attachment K or equivalent procedures - Opening Packages

Attachment L - Air Emissions Questionnaire

Detailed radiation safety procedures for working with lab animals (if applicable)

Detailed radiation safety procedures for human use (if applicable)

- ÿ If this is a new application, have you enclosed the fee?
- ÿ Has the form been signed and dated by corporate management (not the RSO unless the management has filed with the Department a statement authorizing the RSO to sign all applications and radiation safety program commitments?)

RHF-1LI Application for Radioactive Material License Laboratory - Industrial

Washington State Department of Health Division of Radiation Protection

INSTRUCTIONS - Complete all items in this application for a new license or the renewal of an existing license. Use supplemental sheets where necessary. Item 29 must be completed on all applications. Mail in accordance with the directions contained in the application cover letter. Upon review and approval of this application, the applicant will receive a State of Washington Radioactive Material License issued in accordance with the general requirements contained in Washington State Department of Health, Rules and Regulations for Radiation Protection, and the Washington Nuclear Energy and Radiation Control Act, Chapter 70.98 RCW.

Name and Mailing Address of (institution,Firm,Individual own)	of Applicant ner, etc.) INCLUDE ZIP CODE		dress(es) at which Radioa ifferent than Item 1. INCLUI	
2. Person to Contact Regarding	g this Application		Telephone (Area Code -	Number - Extension)
3. This is an Application for: (C	Check appropriate item)	a. New Lice	nse b. Renewal of Lic	ense No
4. Authorized Users (Name of i directly supervise use of Radioact		4A. Training	and Experience (Check or	ne or more)
unecity supervise use of readloads	ive material.)	or	t A completed and attached	
5. Radiation Safety Officer (RS			Radiation Safety Officer (•
designated as Radiation Safety C	onicer.)	Attachmen or	t B completed and attached	d
			duties attached	
Radioactive Material (Elements and Mass number of each.)	7. Chemical and/or Physical Sealed Source Manufactor Model Number		8. Maximum Activity of each Chemical or Source	8A. Maximum Yearly Throughput
a	a		a	a
b	b		b	b
c	c		c	c
				
Device and/or Use Description For Sealed Source, device in the sealed Source in the sealed Source.		lettering in Items	E 6, 7, 8 and 8A above.)Bri	l ief Description of Use -
a.				
b.				
C.				
Attach a complete descriptio per order, activity per use an		n isotope includin	g safety practices used, dis	posal practices, activity

	ection Instruments List -14, S-35, Low energy gan					
MANUFACTURER	MODEL # & TYPE	PROBE MODEL		DETECTS	UNITS	TYPE OF USE
	Detection Instruments (m	andatory for all 1		oassay Program(i		'
instruments pos A. Calibration Free	•		ÿ		will be in accordance wi % of ALI in one year)	th WAC
ÿ Annually and	d after each repair.		ÿ	Attach description	of program, including fre	equency.
•			ÿ	In accordance with	Reg. Guide 8.20 (I125 8	& I131)
B. Calibrations (ch	eck one)		В	oassays will be an	alyzed or done by:	
	ill do calibrations. Attach p t C may be used for dose		ÿ		ompany: List current or	
	will be done by an approve					
-	ency. List current or initial a	-				
Address			ÿ		procedures, including ca	libration of
				analytical equipme	ent.	
License Number	er	1		cilities (all four re		
C. Daily Operationa	I Checks		У		ioactive material are che FM under normal workir made by:	
ÿ List source(s)	used		ÿ	Facilities, equipmen	t, and storage diagram a	attached.
			ÿ	Facilities and equip	ment description attache	ed.
	nitoring (check as approproled monitoring: justification	•	ÿ	Ventilation diagram	or description attached	
	hole Body ÿ Extremi	4	6. S u	rvey Program (che	eck one)	
			ÿ	Attachment D sig	ned and attached, or	
B. Radiation Detec	cted		ÿ	Equivalent survey	criteria and procedures a	attached.
ÿ Beta-Gamm	na ÿ Beta-Gamm				· · · · · · · · · · · · · · · · · · ·	
C. The following do	osimeter requirements will b			-	if applicable) (check one	•
ÿ Dosimeters s	stored in cool, dry place aw		У	• •	act with approved outsic initial or current consulta	
	irces when not in use.					
	at least quarterly; Indicate			Address		
	ÿ Monthly ÿ Quai	-				
Indicate initia	rtified dosimeter supplier wal or current supplier.		ÿ	mailing leak tests	ak tests using approved to kit manufacturer for core	ounting.
Address						
	ng direct reading pocket do edures for use, calibration, o		ÿ	Will do own leak te procedures attache	est including counting. Ded.	etailed

18. Records Management Program

Sample of each record form attached.
 Use Attachment E for list of required forms.

19. Instructions to Personnel

- ÿ Attachment F completed and attached, or
- ÿ Instructions to personnel attached
- ÿ Attachment G Lab Safety Rules completed and attached
- ÿ Equivalent safety rules attached.

20. Waste Disposal

- ÿ Attachment H completed and attached, or
- ÿ Equivalent procedures attached.

21. Emergency Procedures

- ÿ Attachment I completed and attached, or
- ÿ Equivalent procedures attached.

22. Ordering and Receiving Packages

- ÿ Attachment J completed and attached, or
- ÿ Equivalent procedures attached.

23. Opening Packages

- ÿ Attachment K completed and attached, or
- ÿ Equivalent procedures attached.

24. Animal Use

- ÿ Not applicable
- ÿ Detailed radiation safety procedures attached (including waste disposal).

25. Human Use

- ÿ Not applicable
- ÿ Detailed radiation safety procedures attached (including waste disposal).

26. Air Emissions Questionnaire

27. Surety and Decommissioning

ÿ Statement concerning Surety and Decommissioning as defined in WAC 246-235-075 attached.

28. License Fee Required

See WAC 246-254-057

- A. License Fee Category _____
- B. License fee enclosed: \$_____

Item 29 - Certificate (This Item must be completed by Applicant)

The applicant and any official executing this certificate on behalf of the applicant named in Item 1, certify that this application is prepared in conformity with Washington State Department of Health, Radiation Control Regulations and that all information contained herein, including any supplements attached hereto is true and correct to the best of our knowledge and belief.

	By:
(TYPE OR PRINT NAME OF CERTIFYING OFFICIAL)	(SIGNATURE)
(TITLE OF CERTIFYING OFFICIAL)	(DATE)

ATTACHMENT A

Training and Experience Authorized User or Radiation Safety Officer (Use supplemental sheets if necessary.)

1. Name of	f Applicant		(030 30	ppiemental sheets ii necessal	<u>y</u> .)	2. Date of Application	
3. Training Type of	received in basic ra	adioisote Wh	ope handling tech ere Trained	nniques Date(s) & Duration of Tra	inina	Formal Course	On the Job
a. Principle	s and practices on protection		0.0	Jaio(o) a Jaianon oi ma	<u>g</u>	ÿ Yes ÿ No	ÿ Yes ÿ No
standard toring ted instrume						ÿ Yes ÿ No	ÿ Yes ÿ No
use and radioacti	ons basic to the measurement of vity					ÿ Yes ÿ No	ÿ Yes ÿ No
d. Biologica radiation						ÿ Yes ÿ No	ÿ Yes ÿ No
4. Experien Isotope	ce with Radioactive Maximum Amo Used per Exper	unt		of radioisotopes) ience was Gained	D	ate(s) and Duration of Experience	

Updated 1/10/2005 5

Resume' of radiation work experience	ence	
Dates of Employment	Employer Name/Address/ Supervisor/ Phone Number	Job Title/Type of Isotope use
	Supervisor/ Phone Number	••

ATTACHMENT B Duties of the Radiation Safety Officer

- Make sure that all uses of radiation are conducted safely, adhere to the conditions of the license and license application, and result in exposures to personnel which are as low as reasonably achievable (ALARA).
- Act as liaison agent with regulatory authorities, be available for assistance in inspection and audits, and notify the department:
 - A. In writing before making any change which would render the Application for Radioactive Materials License, Radioactive Materials License, Application for Registration (of radiation producing machines, or Notice of Registration no longer accurate.
 - B. Immediately in the event of any radiation accident or incident (including high dosimeter reading):
 - C. Within five (5) days of any positive leak test result of a sealed source; and/or
 - D. Within thirty (30) days in a report stating remedial action taken after accident or incident.
- 3. Be familiar with all applicable regulations and regulatory guides, and assure that license applications are properly filled out and are submitted on time.
- 4. Make sure all surveys, calibrations, and leak tests are performed on time.
- 5. Establish and maintain record systems as applicable for radiation area surveys, wipe tests, leak tests, calibration of instruments, and personnel dosimeter reports.
- 6. If dosimetry is required, advise individual radiation workers of each high exposure report, and conduct a survey to determine the cause of all overexposures so as to preclude reoccurrence. Perform a quarterly review of occupational exposure to authorized users and workers to determine that the exposures are within the limits established for the ALARA program. Annually advise each radiation employee of annual accrued dose including internal and external exposures.
- 7. Assure that individuals working with radiation have appropriate protective devices, including shielding, ventilation, clothing, gloves, remote handling equipment (where necessary), and facilities which aid in keeping exposures As Low As Reasonably Achievable (ALARA).
- 8. Procure and maintain an adequate number of operable and properly calibrated radiation survey instruments, of the appropriate range and type, and properly calibrated counting equipment (if applicant will assay contamination and leak tests).
- 9. Develop and maintain up-to-date operating and emergency procedures.
- 10. Perform, or cause to be performed, semi-annual inventory of all sealed sources received or possessed.
- 11. Post conspicuously "Notice to Employees" RHF-3 and notices of items of noncompliance resulting from department inspections.
- 12. Supply employers of terminated radiation personnel with radiation exposure records as required by regulation.
- 13. Establish and cause to be maintained inventory control of radioisotopes, making sure inventory never exceeds amounts licensed.
- 14. Keep, or cause to be kept, records of receipts of incoming isotopes and surveys of incoming and outgoing shipments.

Attachment B - continued

- 15. Make sure that all incoming and outgoing radioactive shipments are properly packaged and labeled according to DOT requirements, and that shipments are accompanied by proper shipping papers.
- 16. Assure that radioactive materials are disposed of properly, and that records are maintained of all radioactive wastes disposed.
- 17. Perform an annual review of radiation safety program for adherence to ALARA concepts. Make sure that the safety program is followed by all workers dealing with radioactive materials. Investigate any deviation from the program, and take any remedial action necessary.
- 18. Schedule briefings and educational sessions to inform workers of radiation safety rules and procedures:
 - A. For all new personnel,
 - B. With each change in license condition or safety program, and
 - C. Annually in a refresher course for all personnel.

This includes instruction in the ALARA program and philosophy.

- 19. Apprise and inform management of radiation safety status and their responsibilities in maintaining an adequate radiation safety program.
- 20. Take charge in all emergency situations (spills, or release of radioactive material, etc.) to make sure correct emergency decontamination and protection procedures are carried out. Also evaluate the situation that led to the emergency, to reduce the chance of further problems.
- 21. Assure that radioactive materials are used only by individuals authorized by the license.
- 22. Assure that radioactive materials are properly secured against unauthorized removal.

The RSO should have some formal training in radiological health (i.e. college level or its equivalent) and should have specific experience in radiation protection with types, quantities, and use of the radioactive material to be used under the license.

Approved by	Date

ATTACHMENT C CALIBRATION OF EXPOSURE RATE INSTRUMENTS

(Use only if calibrating Dose Rate Instruments yourself.)

- 1. Calibration of survey meters shall be performed with radionuclide sources.
 - A. The Sources shall approximate point sources.
 - B. The source activities or exposure rates at given distances shall be traceable by documented measurements to a standard source certified within five percent accuracy by the U.S. National Bureau of Standards (NBS) calibrations.
 - C. The frequency shall be at least every 12 months and after servicing, except that instruments used for surveys related to industrial radiography shall be calibrated at intervals not to exceed three months and after each servicing and repair.
 - D. Each scale of the instrument shall be calibrated at least at two points located at approximately 1/3 and 2/3 of full scale.
 - E. The exposure rate (mR/hr) measured by the instrument shall differ from the true exposure rate by less than 10 percent at the two points on each scale (read appropriate section of the instrument manual to determine how to make necessary adjustments to bring instrument into calibration.) Readings within ±20 percent will be considered acceptable if a calibration chart, graph, or response factor is prepared, attached to the instrument, and used to interpret meter readings to within 10 percent for radiation protection purposes.
 - F. Records of required calibrations shall be maintained for inspection for a period of at least two years from the date of calibration.

NOTE: Sources of Cs-137, Ra-226, or Co-60* are appropriate for use in calibrations. Since these sources emit rather high-energy photons, they are not suitable for low-energy calibrations that may be required under special circumstances (see Item 3). The activity of the calibration standard should be sufficient to calibrate the survey meters on each scale to be used for radiation protection purposes. Scales up to 1R/hr should be calibrated, but higher-range scales above 1R/hr need not be calibrated when they will not be needed for radiation protection surveys. If there are higher ranges, they should at least be checked for operation and approximately correct response to radiation. Otherwise, a cautionary note that they have not been checked should be placed on the instrument.

These procedures and standards are not appropriate for instruments used to detect or quantify measurements in the I-125 energy range.

- 2. A reference check source of long half-life, e.g., Cs-137 or Ra-226, shall also be read at the time of the above calibration or as soon as the instrument is received from the calibration laboratory. The readings shall be taken with the check source placed in specific geometry relative to the detector. A reading of this reference check source shall be taken:
 - A. Before each use and also after each survey to ensure that the instrument was operational during the survey;
 - B. After each maintenance and/or battery change; and
 - C. At least every three months.

If any reading using the same geometry is not within \pm 20 percent of the reading measured immediately after calibration, the instrument must be re-calibrated (see Item 1.)

^{*} Minimum activities of typical sources are 85 mCi of CS-137, 21 mCi of Co-60, and 34 mCi of Ra-226 (to give at least 700 mR/hr at 20 cm).

Attachment C - continued

Calibration source energies must correspond to energies of radioactive materials to be detected if
instrument response is energy dependent, and if the instrument is to be used for quantitative
measurements in the Xe-133 or Tc-99m energy ranges.

The calibration may be done either:

- As in Item 1 above, with NBS-traceable calibration standards of radionuclides at or near the desired energies, or
- B. As a relative inter-comparison with an energy-independent instrument and un-assumed or uncertified radionuclides.

Alternatively, the manufacturer's energy response curve(s) may be used to correct instrument readings appropriately when lower-energy radiation is monitored.

- 4. Records of the above Items 1; 2-B and -C; and 3 must be maintained.
- Use of Inverse Square Law and Radioactive Decay Law
 - A. An approved calibration source will have a calibration certificate giving its exposure rate at a given distance, or its activity, measured on a specified date by the manufacturer or NBS.
 - (1.) The Inverse Square Law may be used with any point source* to calculate the exposure rate at other distances.
 - (2.) The Radioactive Decay Law may be used to calculate the exposure rates or source activities at times other than the calibration date.
 - B. Inverse Square Law

Consider a point source of radiation at position S, as shown in Figure C-1. Then, the relationship between exposure rates R $_1$, and R $_2$ at detector positions P $_1$ and P $_2$, which are the distances D $_1$ and D $_2$ from S, respectively, is given by the following equation:

$$R_2 = \frac{{D_1}^2}{{D_2}^2} x R_1$$

Where R_1 and R_2 are exposure rates in the same units (e.g., mR/hr, R/hr) and D_1 and D_2 are the distances in Figure C-1 in the same units (e.g., m, cm, ft)

^{*} A source may be considered a point source when the source and the radiation detector are small, in any dimension, compared to the distances at which radiation is to be measured. The center of the detector should be at distances D_1 or D_2 as shown in Figure C-1.

Attachment C - continued

C. Radioactive Decay Law

$$R_t = R_o e^{-\left(\frac{0.693}{T_{1/2}}t\right)}$$

Where

 R_0 and R_t are in the same units (e.g., mR/hr or R/hr)

 $\begin{array}{lll} {\sf R}_{\sf O} & & \text{is exposure rate on the specified calibration date} \\ {\sf R}_{\sf t} & & \text{is exposure rate at some time later designated, } \ {\sf t} \\ {\sf T}_{1/2} \ {\sf and} \ \ {\sf t} & & \text{are in the same units (years, months, days, etc.)} \end{array}$

T_{1/2} is radionuclide half-life

t is the time elapsed between calibration and present time

Example: Source output is given by calibration certificates as 100 mR/hr at one (1) foot on March 10, 1975. Radionuclide half-life is 5.27 years.

Question: What is the output at three (3) feet on March 10, 1977 (2.0 years later)?

(1). Output at 1 foot, 2.0 years after calibration date:

$$-\left(\frac{0.693x\ 2.0}{5.27}\right)$$
 R = 100 mR/hr x e

 $= 100 \times 0.77 = 77 \text{ mR/hr}$ at 1 foot on March 10, 1977.

(2). Out put at three (3) feet, 2.0 years after calibration date:

$$R = \frac{\left(1 ft\right)^2}{\left(3 ft\right)^2} \times 77 \text{ mR/hr}$$

= $\frac{1}{\Omega}$ x 77 = 8.6 mR/hr at 3 feet, 2.0 years after calibration.

Approved by: ______ Date _____

Attachment C - continued Sample Form

Certificate of Instrument Calibration

Instr	ument:			Probe: (if detachable	e)
Manufacturer			Manufacturer	·	
Туре			Type		
Model No			Model No		
Serial No			Serial No		
Calibration Data	ı:				
Scale Ra	Exposure ate (mR/hr)			Adjusted Instrument Reading (mR/hr)	
mments:					
libration Source:					
nufacturer/Model	No		Seria	l No	
clide	Accuracy	Origii	nal Activity/Date	/	
			0	ent Activity	

ATTACHMENT D Standard Survey Program for Unsealed Radioactive Material

- 1. A reference check source of long half-life, e.g., Cs-137 or Ra-226, shall be read and recorded at the time of the initial receipt of the instrument or as soon as the instrument is received from calibration. The readings shall be taken with the check source placed in specific geometry relative to the detector. A reading of this reference check source shall be taken before and after each survey, and after each maintenance and/or battery change. If any reading using the same geometry is not within ± 20 percent of the reading measured immediately after calibration, the instrument must be re-calibrated.
- 2. Use areas shall be surveyed on a weekly basis unless they meet one of the criteria below. Any room meeting this criteria may be surveyed monthly.
 - A room, such as a counting, equipment, or autoradiograph room, where all material is contained and unopened.
 - B. Activity used per month is no more than the values listed in Table I.
 - C. Long term waste storage areas which are accessed infrequently (once a month or less).
- 3. Daily or after use, surveys will be done of clothing, hands and work areas when working with high energy beta, gamma emitters, or I-125, using an appropriate low range instrument. After-use surveys will be done if a spill is suspected.
- Special surveys will be done of work areas immediately after working with quantities greater than or equal to fifty (50) times the values in Table I.
- 5. The weekly, monthly, and special surveys will consist of :
 - A. A direct measurement of radiation levels with a survey meter sufficiently sensitive to detect 0.05 mR/hr or 10,000 cpm. This measurement is not needed for low energy beta emitters.
 - B. A series of wipe tests to measure contamination level. The method for performing wipe tests will be sufficiently sensitive to detect 100 dpm per 100 cm² for the radionuclide involved. If you are using a portable instrument to measure the wipes, go to a low background area to count the wipes.
- 6. A permanent record will be kept of all survey results, including negative results. The record will include:
 - A. Location, date, and person doing the survey.
 - B. Identification of equipment used, including serial number and pertinent counting efficiencies.
 - C. Drawing of the area surveyed, identifying relevant features such as storage areas, waste areas and major use areas. Also include numbered locations corresponding to the places surveyed.
 - D. Measured exposure or count rates, keyed to locations on the drawing.
 - E. Results of contamination survey, keyed to locations on the drawing.
 - F. A background reading for the portable instrument and the results of a blank wipe for non-portable counters.
 - G. Corrective action taken when survey results exceed action levels.

Updated 1/10/2005

ATTACHMENT D Standard Survey Program for Unsealed Radioactive Material continued

7. Action level for direct survey and contamination surveys will be three times the background reading in an area with no radioactive material present or three times the reading on a blank wipe.

	Table I				
Radionuclide	Monthly Throughput				
	(megabecquerels)	(microcuries)			
Hydrogen 3	296	8000			
Carbon 14	7.4	200			
Phosphorus 32	2.22	60			
Phosphorus 33	22.2	600			
Sulfur 35	37	1000			
Calcium 45	7.4	200			
Chromium 51	37	1000			
Cobalt 60	1.85	50			
lodine 125	0.15	4			
lodine 131	0.11	3			
Thorium 228	0.02	0.6			
Uranium 238	0.037	1			
Plutonium 239	0.003	0.08			
Plutonium 241	0.15	4			

For radionuclides not listed, take 1/10 of the lowest of the Annual Limits on Intake listed in WAC 246-221-290 Table I Columns 1 and 2.

Approved by:	Date

ATTACHMENT E Records Required for Laboratory - Industrial License

Laboratory and industrial radioactive material licensees are required to maintain a number of records. This form is designed to simplify your task of complying with the regulations in regard to record keeping. The different record requirements are listed below in the order they appear in WAC 246. It is your responsibility to develop your own records and make sure that all required records are accounted for.

Because the RHF-1LI Application Form is used for a wide variety of laboratory and industrial applications, ranging from small laboratories to large scale industrial uses of radioactive materials, this attachment may list a number of records which may not apply to all users. If a record described in the attachment is for some material or equipment that you will not have, write N/A in the space to the right of the records description. If the need for a particular record would arise only rarely, you need not compose a special form for that record, but may document the required information by letter or memo.

Provide a copy of all records used.

Section of WAC	Type of Records	Form Code No.
246-220-020 (1)	General provisions require records of receipt, use, storage, transfer, and/or disposal of radiation sources.	
246-249-030 (1)	Waste shipment certification to accompany each shipment of radioactive waste to the low-level radioactive waste burial site (may use departmental form RHF-31).	
246-233-020 (4) (c) (iv)	For licensees with generally licensed gauges, maintain records of tests for leakage of radioactive material and proper operation of the on-off mechanism or indicator in compliance with WAC 246-233-020 (4) (c) (ii) and (iii).	
246-221-020 (4)	Records of employee's prior dose history (may use departmental form RHF-4A).	
246-221-080 (2)	Record of leak tests.	
246-221-090 (2) (d)	Record of exposure circumstances of personnel monitoring device when it was not worn.	
246-221-110	Records of surveys preserved as specified in WAC 246-221-230 (this includes receipt surveys, contamination surveys, etc.)	
246-221-190	Records of disposal by release into sanitary sewerage systems proving compliance with limits in WAC 246-221-190.	
246-221-230 (7)	Records of radiation exposures on RHF-5A or all other forms containing all the information required by form RHF-5A.	(this information is usually contained in vendor's film badge exposure reports)
246-222-030	Records of instructions to workers.	

ATTACHMENT F Personnel Training Program

1.	The I	Radiatior provide	n Safety Officer or title	struction shall include, bu	ut is not limited to:			
	A.	Genera	I radioactive materials safety rules	:				
	B.	Personi	nel monitoring program (e.g., use,	exchange, storage, reco	rds, and reports):			
	C.	Radiation	on and contamination survey prog	am:				
	D.	Accider	nt, incident, and emergency proced	lures:				
	E.	Radioa	ctive materials work procedures:					
		(1) (2) (3) (4) (5)	ordering, receipt and opening prostorage: use of radioactive materials: waste packaging and storage: transportation procedures.	cedures:				
	F.	Applical	ble state and federal rules and reg	ulations and license con	ditions.			
2.	shall	provide i	n Safety Officer or (title) instruction to ancillary personnel, s may require them to work in the vi					
	A.	All term	s of the license pertinent to radiati	on safety.				
	B.	Identific	ation of areas where radioactive n	naterial us used or stored	l.			
	C.	Potential hazards associated with radioactive material.						
	D.	Radiological safety procedures appropriate to their respective duties.						
	E.	Pertinent state and federal regulations.						
	F.	Rules a	nd procedures of the license.					
	G.	Obligat	ion to report unsafe conditions to t	he Radiation Safety Offic	er.			
	H.	Appropriate response to emergencies or unsafe conditions.						
	l.	Right to be informed of their radiation exposure and bioassay results.						
	J.	regulati	ns where the licensee has posted ions, and copies of pertinent licens plicable correspondence,) as requ	es and license conditions				
or in th	e vicin	nity of, ra	Officer shall verify that personnel dioactive materials, during annual alations, or the terms of the license	refresher training, and w				
		APPR	OVED BY	<u> </u>	DATE			

ATTACHMENT G General Rules for Safe Use of Radioactive Material

Check appropriate boxes which you will institute, and which apply to your facility.

ÿ	1.	Wear laboratory coats or other protective clothing at all times in areas where dispersible radio- active materials are used.
ÿ	2.	Wear disposable gloves at all times while handling dispersible radioactive materials.

- \ddot{y} 3. Monitor hands and clothing for contamination after each procedure or before leaving the area.
- ÿ 4. Use shielding devices and/or remote tools when working with millicurie or greater quantities of radioactive materials.
- - B. Do not store food, drink, or personal effects with radioactive material (e.g., in refrigerator).
- ÿ 6. Wear personnel monitoring devices (film badge or TLD) at all times while in areas where radioactive materials are used or stored. These devices should be worn at chest or waist level. Personnel monitoring devices when not being worn to monitor occupational exposures should be stored in a designated low background area.
- ÿ 7. Wear TLD finger badges when manipulating millicurie or greater quantities of radioactive materials.
- ÿ 8. Dispose of radioactive waste only in specially designated drains or properly shielded receptacles.
- $\ddot{\mathbf{y}}$ 9. Never pipette by mouth.
- ÿ 10. Survey laboratory work area for contamination after each procedure, or at the end of the day. Decontaminate if necessary.
- Y 11. Confine radioactive solutions in covered containers plainly identified and labeled with name of compound, radionuclide, date, activity, and radiation level, if applicable.
- ÿ 12. Always transport radioactive material in shielded containers.
- ÿ 13. Use remote tools when handling sealed sources.
- ÿ 14. Leak test nuclear gauges with gauges in the locked, stored, or safe position, but ONLY if the license authorizes you to take the leak tests.

APPROVED BY	DATE

ATTACHMENT H Waste Disposal

1.	Liqui	d waste (other than liquid so	cintillation cocktail) will be disposed or	f (check as appropriate)
	ÿ ÿ ÿ	Solidified and disposed by	em in accordance with WAC 246-221 commercial waste disposal service,	complete Item 4 below.
	ÿ	area with a low-level surv	ey meter and with all shielding remove	els, as measured in a low background yed, have reached background levels. te will be disposed of in normal trash.
2.	Liqui	d Scintillation Cocktail will b	e disposed of:	
	ÿ	biodegradable by Washing Protection Agency.	m in accordance with WAC 246-221-7 gton State Regulations which are mor	
	ÿ	By commercial waste dispo	osal service	
3.	Othe	r solid waste will be dispose	d of (check as appropriate)	AME
	ÿ	By Return to the manufact	urer	
	ÿ	Held for decay for at least area with a low-level surve	ten (10) half lives, until radiation leve ey meter and with all shielding remov	els, as measured in a low background yed, have reached background levels. te will be disposed of in normal trash.
	ÿ	By in-house compaction,	describe under Item 15 of foregoing	g application.
	ÿ	·	osal service, complete Item 4 below	
	ÿ		, ' '	
4.	Th	e commercial waste disposa	al service used will be	
				NAME
	AD	DRESS	CITY	STATE
	Ra	dioactive Materials License	Number	
_	_			
5.	Sa	nitary sewer radioactive ma	terial disposal concentration calculati	ion.
	A.	Determine total volume of	sewage per month	ml
			sewage may be estimated by averaging tater used by a facility as stated on a water	
		Useful Conversions:	1 cubic foot = $2.832 \times 10^4 \text{ ml}$	1 gallon = $3.78 \times 10^3 \text{ ml}$

Attachment H - continued

2) 3)		<u></u>	,, \	per month)	
3)					
,					
41					
,				_	
or each isc	otope, divide t	the activity (mic	rocuries) by the monthl	y volume (ml).	
1	Isotope		Activity + Monthly V	olume = Month	y Concentra
1)		<u></u>	μCi ÷	ml =	μCi/ml
2)		<u>—</u>	μCi ÷	ml =	μCi/ml
3)		<u>—</u>	μCi ÷	ml =	μCi/ml
4)		<u>—</u>	μCi ÷	ml =	μCi/ml
5)			μCi ÷	ml =	μCi/mI

DATE

Updated 1/10/2005 19

APPROVED BY

ATTACHMENT I Emergency Procedures

Minor Spills

- 1. NOTIFY Notify persons in the area that a spill has occurred.
- PREVENT THE SPREAD Cover the spill with absorbent paper.
- CLEAN UP Use disposable gloves and remote handling tongs. Carefully fold the absorbent paper and pad. Insert into a plastic bag and dispose in the radioactive waste container. Also insert into the plastic bag all other contaminated materials such as disposable gloves.
- 4. SURVEY with a low-range, thin-window G-M survey meter, checking the area around the spill, hands, and clothing for contamination.
- 5. REPORT Report incident to the Radiation Safety Officer.

Major Spills

- 1. CLEAR THE AREA Notify all persons not involved in the spill to vacate the room.
- 2. PREVENT THE SPREAD Cover the spill with absorbent pads, but do not attempt to clean it up. Confine the movement of all potentially contaminated personnel to prevent any spreading.
- 3. SHIELD THE SOURCE If possible, the spill should be shielded, but only if it can be done without further contamination or without significantly increasing your radiation exposure.
- 4. SECURE THE ROOM Leave the room and lock the door(s) to prevent entry.
- 5. CALL FOR HELP Notify the Radiation Safety Officer immediately.
- 6. PERSONNEL DECONTAMINATION Contaminated clothing should be removed and stored for further evaluation by the Radiation Safety Officer. If the spill is on the skin, flush thoroughly and then wash with mild soap and lukewarm water.

Radiation Safe	ty Officer _			
Office Phone		 	 	
Home Phone				

Loss, Theft, Fire, Explosion, or Vehicle Accident

Follow the procedures outlined in the Washington State Department of Health - Radiation Emergency Handbook. Principally this shall include:

1. Secure the area around the accident. Keep unauthorized persons away. Alert people in the vicinity as to the presence of radioactivity and a possible hazard.

Attachment I - continued

2.	DC	NOT leave the site. Send a helper or onlooker to notify the following:
	a.	Radiation Safety Officer
		Work Phone Home Phone
	b.	Local Police
	C.	Local Fire Department, where applicable
3.		e Radiation Safety Officer, in turn, must immediately notify the State of Washington Radiation nergency Response at (206) 682 - 5327, which is
		206 N-U-C-L-E-A-R
	an	d other local authorities, as indicated.
4.	hel rac wo	e radiation worker should inform emergency workers of the radiation hazard possibly existing, should be them keep the area secure, and should explain to the emergency personnel the location of the dioactive device or chemical and the extent of the possible hazard. In no case should the radiation or relative the site until qualified experts arrive, unless, of course, said worker is seriously injured or capacitated, and must be removed from the site by emergency personnel.
		Alternate Names and Telephone Numbers Designated by Radiation Safety Officer
		APPROVED BY DATE

ATTACHMENT J Procedures for Ordering and Accepting Delivery of Radioactive Material

- 1. The Radiation Safety Officer, or his/her designee, will place all orders for radioactive materials and will ensure that the requested materials and quantities are authorized by the license and that possession limits are not exceeded.
- 2. A system for ordering and receiving radioactive materials will be established and maintained. The system will consist minimally of the following:
 - A. Ordering of routinely used materials -
 - (1.) Written records that identify the isotope, compound, activity levels, and supplier, etc., will be used.
 - (2.) The written records will be referenced when opening and/or storing radioactive shipments.
 - B. Ordering of specially used materials -
 - (1.) A written request will be obtained from the authorized user who will perform the procedure.
 - (2.) Persons ordering the materials will reference the authorized user's written request when placing the order. The authorized user's request will indicate isotope, compound, activity level, etc.
 - (3.) The authorized user's written request will be referenced when receiving, opening, and/or storing the radioactive material.
 - C. Maintain written records for all ordering and receipt procedures.
- 3. During normal working hours, carriers will be instructed to deliver all radioactive package(s) directly to the nuclear laboratory or to some other official receiving area for radioactive materials.
- 4. During off-duty hours, security personnel or other designated individuals will accept delivery of radioactive package(s) in accordance with the procedures as stated in your facility memorandum as attached.

APPROVED BY	DATE

Attachment J - continued EXAMPLE

To:	Security Personnel
From:	Administrator
Subject:	Receipt of Package(s) Containing Radioactive Material
on Sund nuclear I	kage(s) containing radioactive material that arrive between 4:30 p.m. and 7:00 a.m., or ays shall be signed for by the security guard on duty, and be taken immediately to the laboratory. Unlock the door, place the package on top of the counter immediately to of the door, re-lock the door.
Safety C	ckage is wet or appears to be damaged, IMMEDIATELY contact the company Radiation Officer. Ask the carrier to remain until it can be determined that neither the driver nor ery vehicle is contaminated.
Radiatio	n Safety Officer
Office Ph	none
Home Pl	hone
	Submit a copy of your own company's memorandum
	Calcinia a copy of your company o momeranaam
	APPROVED BY DATE

ATTACHMENT K

Procedures for Safely Opening Packages Containing Radioactive Material

- Special requirements will be followed for packages containing quantities of radioactive material in excess of the Type A, quantity limits (these are packages containing "X" amounts.) They will be monitored for surface contamination and external radiation levels within three (3) hours after receipt if received during working hours or within eighteen (18) hours if received after working hours, in accordance with the requirements of WAC 246-221-160. All shipments of liquids greater than exempt quantities will be tested for leakage. The Department will be notified in accordance with the regulations if removable contamination exceeds 0.01 μCi/100 cm or if external radiation levels exceed 200 mR/hr at the package surface or 10 mR/hr at three (3) feet (or 1m.)
- 2. For all packages, the following additional procedures for opening packages will be carried out:
 - A. Put on gloves to prevent hand contamination.
 - B. Visually inspect package for any sign of damage (e.g., wetness, crushed, etc.) If damage is noted, stop procedure and notify Radiation Safety Officer.
 - C. Measure contamination on external surface of any package labeled with a radioactive white I, yellow II, or yellow III label, unless material is in gas form, or a special form as delineated on the packaging slip.
 - D. Measure surface count or rate and record. If reading is greater than _____mR/hr, or ____ CPM (insert appropriate action levels for your facility,) stop procedure and notify Radiation Safety Officer.
 - E. Open the package with the following precautionary steps:
 - (1.) Open the outer package (following manufacturer's directions, if supplied,) and remove packing slip.
 - (2.) Open the inner package and verify that contents agree with those on packing slip. Compare the requisition, packing slip, and label on container.
 - (3.) Check integrity of final source container (i.e., inspect for breakage of seals or vials, loss of liquid, and discoloration of packaging material.)
 - (4.) Check also that receipt of this shipment does not exceed possession limits.
 - F. Monitor the packing material and package(s) for contamination before discarding.
 - (1.) If contaminated, treat as radioactive waste.
 - (2.) If not contaminated, obliterate radiation labels before discarding in regular trash.
- 3. Maintain records of the results of checking each package, using "Radioactive Shipment Receipt Record" (see next page), or a form containing the same information.

APPROVED BY DATE

Type A Package Quantity Limits for Selected Radionuclides

Radionuclide Padionuclide	(Unsealed Material)	<u>Ci</u>
H-3		1000
Carbon 14		60
Sulfur 35		10
Phosphorus 32		60
Cesium 137		30
lodine 125		70
235 Uranium (92)		0.2
226 Radium (88)		0.05

Attachment K - continued EXAMPLE

Radioactive Shipment Receipt Report

1.	PO No	Survey Da	ate	
	Time	Surveyor		
2.	Condition of Package: ÿ OK	ÿ Punctured	ÿ Wet	ÿ Crushed
	ÿ Other			
3.	Radiation Units on Label			
4.	Measured Radiation Levels			
	Instrument Used		_	
	Background		_	
	Response Check		_	
	Package Surface		_ mR/hr or CPM	
5.	Do Packing Slip and Vial Contents Ag	gree?		
	A. Radionuclide ÿ Yes ÿ N	No, difference _		
	B. Amount ÿ Yes ÿ N	No, difference _		
	C. Chem. Form ÿ Yes ÿ N	No, difference _		
6.	Wipe Results From:			
	Outer CPM =		_ DPM	
	eff = ()			
7.	Survey Results of Packing Material and If H-3 or C-14, a wipe of inner	· · · · · · · · · · · · · · · · · · ·		CPM
8.	Disposition of Package After Inspection	on		
9.	If Department and/or Carrier Notificat notified.	ion Required, C	Give Time, Date, a	nd Persons

Updated 1/10/2005 25

If Radiation Safety Officer notification is required, he/she must sign below.

10.

ATTACHMENT L

Air Emissions Questionnaire for Exempt Status

(Complete table below for Items 1, 2, and 3.)

H-3

H-3

1. List all radionuclides to be authorized on your license. Use additional sheets as necessary.

1 Ci per year

2 Ci per vear

- Estimate the Annual Possession Quantity, the sum of the quantity of a radionuclide possessed at the beginning of
 the calendar year plus the quantity of radionuclide received or produced during the calendar year. Perform this
 calculation for each radionuclide listed in Item 1.
- 3. For each radionuclide list all physical forms that may be exhibited during possession. If a radionuclide will exist in more than one form, give separate listings for each. List the most restrictive physical form (gas, liquid, then solid, capsules are considered solid) that any quantity of radionuclide will exhibit. For instance, if you will use 2 curies of tritiated water to produced H-3 gas and 1 curie of tritiated water that will remain in the aqueous phase your response should be:

liquid

gas

1.	Radionuclide	2.	Annual Possession Quantity (curies/year)	3.	Physical Form (gas, liquid, or solid)
Α.		A.		Α.	
B.		B.		B.	
C.		C.		C.	
D.		D.		D.	
E.		E.		E.	

Sign	eature/ Title:	Da	ite:			
Facil	ity name: Facility Address:					
13.	Nearest individual/farm producing meat					meters
12.	Nearest individual/farm producing milk					meters
11.	Nearest individual/farm producing vegetables					meters
the d	istance in Question 7 should be recorded.)					
	ance from the maximally exposed individual to: getables, milk or meat for individual consumption are produced at the location of the m	axin	nally e	xpos	ed ir	ndividual
10.	Stack Volumetric Flow Rate					m ³ /s
9.	Stack Diameter	_				meters
If Qu	estion 8 is "Yes" complete Questions 9 and 10.					
8.	Is the maximally exposed individual in the same building?		Yes		No	
7.	Distance to Nearest Residence/Business (Maximally Exposed Individual)					meters
6.	Building Width					meters
5.	Building Height					meters
4.	Release Height (the distance from ground level to the top of the stack or release point)					meters

TO: New and Renewal License Applicants

SUBJECT: CITY/COUNTY OFFICIAL NOTIFICATIONS

The Department of Health has adopted a local Government Notification Policy as required by law. The Division of Radiation Protection notifies the appropriate local government officials whenever a Radioactive Materials License application is received. In order to allow local officials time to comment if they wish, the Division of Radiation Protection is required to delay the issuance of licenses for at least 20 days from the date of such notification. In order to expedite this notification process, please supply the information requested below.

NOTE: If you have more than one facility, your answers must pertain to the facility/facilities where radioactive material <u>is to be used and/or stored</u>.

RETURN THIS DOCUMENT WITH YOUR RADIOACTIVE MATERIALS LICENSE APPLICATION.

N	Name of City:
1	Name of Mayor:
(Complete Mailing Address of Mayor (Street, P.O. Box, City, State, Zip Code):
_	
- JC	UR FACILITY IS LOCATED <u>OUTSIDE</u> THE CITY LIMITS, PLEASE INDICATE
ľ	Name of County:
1	Name of County Commissioner:
1	Name of County Commissioner: